Docusign Envelope ID: 63DD68EF-BB39-83C4-80A5-9B42B43CA94D

2 Tech Drive
Andover, MA 01810
www.mksinst.com

UK Declaration of Conformity

Application of Council Directive(s):

☒ Electromagnetic Compatibility Directive (EMCD) – 2014/30/EU
☒ Restriction of Hazardous Substances Directive (RoHS3) – (EU) 2015/863(7)
☒ Waste Electrical and Electronic Equipment – Directive 2012/19/EU

Standard(s) to which conformity is declared:

☒BS EN 61326-1:2013 – (EMC)
☒BS EN ISO 12100:2010 Safety of Machinery – General Principles of Design – Risk Assessment and Risk Reduction

Emissions:

☒EN 55011: 2016 +A1:2017 (4) Industrial, scientific and medical equipment - Radio-frequency disturbance characteristics - Limits and methods of measurement

Immunity:

☒BS EN 61000-4-2:2009 EMC/Electrostatic Discharge Immunity Test
☒BS EN 61000-4-3:2006+A2:2010 EMC/Radiated Radio Frequency Electromagnetic Field Immunity Test
☒BS EN 61000-4-4:2012 EMC/Electrical Fast Transient/Burst Immunity Test
☒BS EN 61000-4-5:2014+A1:2017 EMC/Surge Immunity Test
☒BS EN 61000-4-6:2014 EMC/Conducted Disturbances induced by Radio Frequency Fields Immunity Test

Manufacturers Name: MKS Instruments, Inc., 2 Tech Drive, Andover, MA 01810 USA

Authorized Representatives Name & Location: MICRO-CONTROLE Spectra-Physics, 9 rue du Bois Sauvage
91055 Evry FRANCE

Equipment Type/Description: **Rotary stage with stepper motor**

Model Number(s) (6): **FCR100**

The object of the declaration described above is in conformity with the relevant Community harmonization legislation. MKS product conforms to the above Directive(s) and Standard(s) only when installed in accordance with manufacturer's specifications. This declaration has been issued under the sole responsibility of the manufacturer.

Date: 10/23/2026

Signature : Signé par : jean-louis nicque 9FE1AFCFA2EB4E7...

Full Name : **Jean-Louis NICQUE**

Title : **General Manager**

---

4) Class A, Group 2
6) Compliance of the above model numbers requires the use of a braided shielded cable properly terminated at both ends – if so noted in the MKS Instruction Manual.
7) RoHS Directive has to be checked for in scope products; cannot CE mark without compliance to RoHS. RoHS Directive can be unchecked only for systems which MKS sells which qualify for "Large Scale Industrial Tool" exclusion.